



# BAKERBOND™ XWP 500

## PolyCSX-35

35um, 500Å  
Hydrophilic Polymer-CSX

Product No. 7587  
Lot No. L14088  
Release Date 04/19/2012

### Certificate of Analysis

TEST	SPECIFICATION	RESULT
Appearance	Passes Test	Passes Test
Mean Particle Size um	30 - 40	37
Carbon (C)	Information Only %	50 %
Nitrogen (N)	3.5 - 5.5 %	5.0 %
Sulfur (S)	2.0 % min.	2.5 %
Pressure (psi)	Passes Test	Passes Test
Liquid Chromatography Use Test	Passes Test	Passes Test
Glucose Oxidase	Information Only	1.50
IgG, K'	Information Only	3.9
Lysozyme, K'	Information Only	6.2
g Globulin Capacity (by Frontal Chromatography)	Information Only	50
Bacterial Count	Information Only	< 1

For Laboratory, Research or Manufacturing Use

Store between 4°C and 15°C

Country of Origin: USA



Phillipsburg, NJ 9001.2008, 14001.2004  
Paris, KY 9001.2008  
Mexico City, Mexico 9001.2008  
Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003  
Salangor, Malaysia 9001.2008  
Panaji, India 9001.2008  
Gliwice, Poland 9001.2008, 17025.2005

Richard M. Siberski  
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor™ Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610