



BAKERBOND™ XWP 500

Poly PEI-35

35um, 500Å
Polyethyleneimine Bonded on
Hydrophilic Polymer

Product No. 7585
Lot No. K12085
Release Date 04/13/2011

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Appearance	Passes Test	Passes Test
Mean Particle Size um	30 - 40	36
Carbon (C)	Information Only %	4.0 %
Nitrogen (N)	4.5 - 6.5 %	5.8 %
Pressure (psi)	Passes Test	Passes Test
Liquid Chromatography Use Test	Passes Test	Passes Test
Lysozyme (Retention time, min)	Information Only	1.40
IgG, K'	Information Only	2.2
BSA, K'	Information Only	6.3
BSA Capacity	Information Only	27
Bacterial Count	Information Only	1

For Laboratory, Research or Manufacturing Use

Store between 4°C and 15°C

Country of Origin: USA

ISO

Phillipsburg, NJ 9001.2008, 14001.2004
Paris, KY 9001.2008
Mexico City, Mexico 9001.2008
Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
Selangor, Malaysia 9001.2008
Panaji, India 9001.2008
Gliwice, Poland 9001.2008, 17025.2005

Marcy M. Malloz

Marcy M. Malloz
Director of QA & Regulatory Affairs

For questions on this Certificate of Analysis please contact Technical Services at 1-800-582-2537 or 908-859-2151

Avantor™ Performance Materials, Inc. (formerly Mallinckrodt Baker, Inc.)

222 Red School Lane • Phillipsburg, NJ 08865 • Phone: 908.859.2151 • Fax: 908.859.6905